

BR 9400 ASTM Distillation Columns



These columns are designed to separate fuel components to determine fuel quality as defined by ASTM standard methods D2892 and D5236. The system is semiautomatic with microprocessor control. The vacuum level, bath temperature and heating rates are programmable, with up to 50 stored programs. The system will process 1-liter batches and separate crude oils into 8 fractions.

This instrument can be used to determine oil quality as defined by the two ASTM standard methods. It is currently being used to characterize TDO oil, as well as other biofuels, such as upgraded pyrolysis oil.

This new column was purchased with funds from a TRC construction and equipment grant provided by MTAf.